


Environmental Noise Level Check Procedure

Purpose: To ensure the room is quiet enough for screening. Person with normal hearing should be able to hear 10 dB below “passing”.

Perform Pure Tone Audiometry Screening twice.

- Start with Right Ear
 - 10 dB at 1000, 2000, 4000 and 6000 Hz
- Switch to Left Ear
 - 10 dB at 6000, 4000, 2000, 1000 Hz
 - Increase to 15dB at 500 Hz
- Switch to Right Ear
 - 15 dB at 500 HZ

Volume (dB)	15	10	10	10	10
Frequency (Hz)	500	1000	2000	4000	6000
Right Ear	Finish 10	Start 1	2	3	4
Left Ear	9	8	7	6	5



Minnesota Department of Health
 Child and Teen Checkups
 651-201-3650
health.childteencheckups@state.mn.us
www.health.state.mn.us

12/2022

To obtain this information in a different format, call: 651-201-3650.